CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	Antithrombin 10 x 10 IU	PRODUCT # :	82072039
EXP. DATE:	2019-11	LOT # :	N0570188
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	654023

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

KIT COMPONENTS :

NAME	LOT NUMBER	EXP.DATE	VOLUME
ANTITHROMBIN 10 IU	N0570188	2019-11	1 ml

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV1 and 2, and antibodies to HCV.

PREPARED BY: Maria Oettinger DATE: 6-22-17

APPROVAL SIGNATURE: P. Cenega DATE: 6/22/17

Form QCM 654023-2, Rev 01/09

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000