

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	COATEST APC Resistance Control Plasma Level 1	PRODUCT # :	82265063
EXP. DATE:	2020-02	LOT # :	N0285768
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	C00423

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R:		MIN	MAX	UNITS		PASS/FAIL
APT Time	CaCl ₂	28	34	seconds	32	PASS
	APC/CaCl ₂	90	--	seconds	117	PASS
APC Ratio		3.5	4.4	--	3.7	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	--	38	seconds	36	PASS
	APC/CaCl ₂	75	--	seconds	112	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

Kit Components:

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 1	N0285768	2020-02	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel N.M. Patel

DATE: 2/27/2018 2-27-18

APPROVAL SIGNATURE: D. Tencza D. Tencza

DATE: 2/27/2018 2/27/18