

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

PRODUCT:	COATEST APC Resistance Control Plasma Level 2	PRODUCT # :	82266863
EXP. DATE:	2020-01	LOT # :	N0185245
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	H01444
		DATE OF MANUFACTURE:	1/30/2018

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			
<b>Analysis with COATEST APC R:</b>						
APT Time	<u>CaCl<sub>2</sub></u>	27	44	seconds	40	PASS
	<u>APC/CaCl<sub>2</sub></u>	--	101	seconds	74	PASS
APC Ratio		--	2.3	--	1.8	PASS

<b>Analysis with COATEST APC R V:</b>						
APT Time	<u>CaCl<sub>2</sub></u>	27	40	seconds	37	PASS
APC Ratio		1.4	1.9	--	1.7	PASS

**Kit Components:**

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
Control Plasma Level 2	N0185245	2020-01	1 ml (x 5)

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: M. Oettinger *M. Oettinger*

DATE: 2/27/2018 *2-27-18*

APPROVAL SIGNATURE: D. Tencza *D. Tencza*

DATE: 2/27/2018 *2/27/18*