

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	<u>COAMATIC FACTOR VIII</u>	EXPIRATION DATE: <u>2020-01</u>
PRODUCT NUMBER :	<u>82258563</u>	
LOT NUMBER:	<u>N0184990</u>	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.43	PASS
Intercept	---	0.12	--	0.05	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.6	PASS
Intercept	---	0.21	--	0.10	PASS
r	0.990	---	--	0.999	PASS

We hereby certify that the above product(s):
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel N.M. Patel
 PRINT SIGNATURE

DATE: 3-5-18
3/5/2018

TITLE: Senior QC Analyst
 APPROVAL SIGNATURE: Donna Tencza Donna Tencza
 PRINT SIGNATURE

DATE: 3/5/18
3/5/2018

TITLE: Senior QC Analyst