



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7BM0AZ

TECHNOZYM[®] PAP ELISA Control Set 2x0.5mL

REF TC12064

LOT 7BM61B0.01

 31-Jul-2018

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x 0.5mL	PAP High Control	ABM61B0	31-Jul-2018
1 x 0.5mL	PAP Low Control	BBM61B0	31-Jul-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (ABM61B0)	98.9 – 164.9 ng/mL	128.4	OK
	Low Control (BBM61B0)	13.0 – 30.2 ng/mL	19.8	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

12. OKT. 2016

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

12. OKT. 2016

Datum / Date: