

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ8F00AZ
---	---	----------

## TECHNOTHROMBIN® TGA SUB 5x1.5mL

REF    5006235  
LOT    8F71B00.02  
    2019-04-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1.5mL	TGA Substrat	8F71B00	2019-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Z72B00	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F72B00	375 - 625	520.4	OK
		TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D62B01	251 - 418	326.3	OK
	TGA Control Low (nM Peak Thrombin)	TGA RC High <span style="border: 1px solid black; padding: 2px;">LOT</span> 7F72B00	257 - 429	346.1	OK
	<span style="border: 1px solid black; padding: 2px;">LOT</span> 2881B00	TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D62B01	132 - 220	197.4	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

**02. FEB. 2018**

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *Michael Kafka*

**02. FEB. 2018**

Datum / Date: