



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3U00AZ.01

TECHNOVIEW Arixtra Control L 6x1mL

REF 5090012

LOT 3U71B00.02

🕒 2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.32 – 0.59 µg/mL	0.45	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.32 – 0.59 µg/mL	0.41	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Suzana GRESAK PODAKAR
Suzana Gresak Podakar

12. MRZ. 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER
Eva Maria Hager

12. MRZ. 2018

Datum / Date: