

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1R00.01
---	---	-----------

**Coagulation Reference, 50x1mL**

REF      5220120  
LOT      1R71B00.01  
      28-Feb-2019

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen Clauss method, diluted plasma	2.70 – 4.06 g/L	3.70	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Lucas ROSENMAYR**  


---

12. JULI 2017  


---

 Datum / Date:

**Freigabe von  
Approved by**

  
**Michael KAFKA**  


---

13. JULI 2017  


---

 Datum / Date: