



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0800AZ.01

TECHNOCHROM<sup>®</sup> anti-Xa Kit 80T.

REF 5340250

LOT 0872B00.02

⌚ 2019-08-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x20 mL	R1 anti Xa Buffer	9R71B00	2022-08-31
1x4 mL	R2 FXa	7K71B00	2019-08-31
1x4 mL	R3 Substrat	8A71B00	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Anti-Xa Rivaroxaban Ceveron <sup>®</sup> alpha	Rivaroxaban Cont M LOT 3571B00	86.2 – 143.6 ng/mL	110.4	OK
	Rivaroxaban Cont H LOT 3371B00	242.0 – 403.4 ng/mL	307.8	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

04. DEZ. 2017  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

04. DEZ. 2017  
Datum / Date: