



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRK00AZ

TECHNOZYM® vWF:CBA ELISA Collagen Typ I Kit 96T

REF 5450311

LOT RK81B00.01

2019-03-31

Inhalt Content	Reagenz Reagent	LOT	
1 x	vWF:CBA Plate Collagen Typ I	RNM71B0	2019-03-31
1 x 0.3mL	vWF:Ag Conjugate	RNH61B1	2019-12-31
1x 80mL	Wash buffer Concentrate	RNA72B0	2019-11-30
1 x 90mL	Incubation buffer	RNB73B0	2019-10-31
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	vWF:CBA Calibrator 1	1ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 2	2ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 3	3ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 4	4ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Calibrator 5	5ND71B0	2019-12-31
1 x 0.5mL	vWF:CBA High Control	AND71B0	2019-12-31
1 x 0.5mL	vWF:CBA Low Control	BND71B0	2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.49	OK
	Calibrator 5	< 0.1	0.07	OK
Verwendbarkeit Usability	High Control (AND71B0)	0.72 – 1.20 IU/mL	0.83	OK
	Low Control (BND71B0)	0.17 – 0.41 IU/mL	0.28	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

28. MRZ. 2018

Datum / Date:

**Freigabe von
Approved by**

Eva Maria HAGER
Eva Maria HAGER

28. MRZ. 2018

Datum / Date: