	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ9600.01
---	---	-----------

**Ceveron® Cleaning Solution 3% 25 mL**

REF      9820300  
LOT      9681B00.01  
      2019-03-31

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.1	OK
Volumen Volume	25.0 – 26.5 mL	25.0	OK
Sterilität Sterility	0 CFU/mL	0	OK


Verwendbarkeit, Fibrinogen, Clauss method Ceveron® alpha Usability, Fibrinogen, Clauss method Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 2px;">LOT</span> 1P71B00	3.17 – 4.75 g/L	4.11	OK
	CCA <span style="border: 1px solid black; padding: 2px;">LOT</span> 3P71B00	1.02 – 1.54 g/L	1.23	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
**Eva Maria HAGER**

10. APR. 2018  
Datum / Date:

Freigabe von  
Approved by

  
**Suzana GRESAK PODAKAR**

10. APR. 2018  
Datum / Date: