

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	Antithrombin 10 x 10 IU	EXPIRATION DATE:	2020-06
PRODUCT NUMBER:	82072039		
LOT NUMBER:	N1274643		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
ANTITHROMBIN, 10 IU (654023)					
ANTITHROMBIN ACTIVITY	9	11	IU/vial	10	PASS

All human source material used to manufacture this product have been tested according to FDA guidelines and found to be non-reactive for: HBsAg, antibodies to HIV 1 and 2, and antibodies to HCV.

3-14-18

PREPARED BY: N.Patel *p.m. Patel* DATE: 3/14/2018
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: D.Tencza *D. Tencza* DATE: 3/14/2018 *3/14/18*
 PRINT SIGNATURE

TITLE: Senior QC Analyst