

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance V-S	EXPIRATION DATE: 2019-08
PRODUCT NUMBER:	82313863	
LOT:	N0286090	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	Pass/Fail
	MIN	MAX			
APT-time on ACL					
Diluted Control Plasma Level 1	30	47	seconds	33	PASS
Diluted Control Plasma Level 2	31	51	seconds	34	PASS

APC-V ratio on ACL					
Control Plasma Level 1	2.8	3.2	--	2.9	PASS
Control Plasma Level 2	1.4	1.9	--	1.7	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N.m. Patel*
PRINT SIGNATURE

DATE: 3/26/2018 *3-26-18*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *D. Tencza*
PRINT SIGNATURE

DATE: 3/26/2018 *3/26/18*

TITLE: Senior QC Analyst