



Certificate of Analysis

PRODUCT:	CoaTest SP Factor FVIII Kit	PRODUCT #:	82408663
EXP. DATE:	2019-11	LOT #:	N1174050
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	824086

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
Coatest SP Factor VIII Kit (82103363)					
High Range					
Slope	0.0077	0.0112	--	0.0084	PASS
Intercept	--	0.250	--	0.105	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
Low Range					
Slope	0.020	0.060	--	0.039	PASS
Intercept	--	0.406	--	0.125	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

KIT COMPONENTS :			
NAME	LOT NUMBER	EXP. DATE	VOLUME
CaCl ₂	N1174040	2020-11	6 ml
CT SP Substrate	N1174042	2019-11	12 ml
Factor Reagent	N1174049	2020-05	10 ml
CT SP Buffer, pH 7.3	N1173752	2019-11	20 ml
CT SP Phospholipid	N1173753	2019-11	2 ml

PREPARED BY: TRATEL N.M. Patel DATE: 12/19/17 12-19-17

APPROVAL SIGNATURE: DTENCZA D. Encza DATE: 12/19/17 12/19/17