



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ...00AZ

TECHNOZYM® PAP Complex ELISA Kit 96T.

REF TC12060

LOT 120603B7.01

⌚ 2019-02-28

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	PAP Plate	RGY71B0	2019-04-30
1 x 0.3mL	anti PAP POX	RBK61B1	2019-09-30
2 x 20mL	Wash buffer Concentrate	RBE71B0	2019-02-28
1 x 20mL	Dilution Buffer	RAF71B0	2019-11-30
1 x 12mL	TMB Substrate	RKN71B0	2019-05-31
1 x 12mL	Stop solution	RKK71B0	2019-05-31
1 x 0.5mL	PAP-Calibrator 1	1BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 2	2BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 3	3BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 4	4BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 5	5BM61B1	2019-09-30
1 x 0.5mL	PAP-high Control	ABM61B1	2019-09-30
1 x 0.5mL	PAP-low Control	BBM61B1	2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.41	OK
	Calibrator 5	OD: < 0.2	0.02	OK
Verwendbarkeit Usability	High Control (ABM61B1)	98.9 – 164.9 ng/mL	123.58	OK
	Low Control (BBM61B1)	13.0 – 30.2 ng/mL	21.73	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

30. NOV. 2017

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

30. NOV. 2017

Datum / Date: