



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6C0A.03

TECHNOCLOT® PT PLUS 5x4mL

REF 5003220

LOT 6C77BA0.01

🕒 2019-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement		Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN	PT Plus (%)	68.7 – 86.5	78.0	OK
	LOT 1P71B00	PT Plus (INR)	0.93 – 1.17	1.06	OK
	CCA	PT Plus (%)	10.9 – 13.7	11.7	OK
	LOT 3P71B00	PT Plus (INR)	2.59 – 3.27	3.03	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von  
Prepared by

  
Stefan STEINER

09. MRZ. 2018  
Datum / Date:

Freigabe von  
Approved by

  
Eva Maria HAGER

09. MRZ. 2018  
Datum / Date: