
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1A00.02
---	---	-----------

Technoclot Control N 10x1 mL

REF	5020070
LOT	1A71B01.01
	2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Siron LS (sec)	21.9 – 27.1	25.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Christoph ENNE

07. DEZ. 2017
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

07. DEZ. 2017
Datum / Date: