



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQP000.02

Att. 18

P781B00.01

## Technoview® Edoxaban CONT Low 5x1mL

REF	5090251
LOT	P781B00.01
	2020-03-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Cap color	red confirmed OK
	Vial type	clear glass vial confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Batch label	present confirmed OK
	Appearance before reconstitution	buff colored plug of lyophilized material confirmed OK
	Appearance after reconstitution	buff colored liquid confirmed OK
Virologie / Virology  The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg Negative HIV-Ak / HIV-Ab Negative HCV-AK / HCV-Ab Negative	Negative Negative Negative OK
Kontrollwiederfindung am Ceveron® Alpha Recovery of controls on Ceveron® Alpha	Technochrom® anti-Xa LOT 0872B00 21.0 - 39.0 ng/mL	29.1 ng/mL OK

Durchgeführt von  
Prepared by

**Christoph ENNE**

30. MRZ. 2018

Datum / Date:

Freigabe von  
Approved by

**Stefan STEINER**

30. MRZ. 2018

Datum / Date: