

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM® FXIII 3 x 3 mL

REF 5360010
LOT 0Z83B00.01
 2019-04-30

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2305100/1	2019-04-30
3 x 3 mL	FXIII Detection Reagent	2305100/2	2019-04-30
3 x 3 mL	FXIII NADPH	2305100/3	2019-04-30
3 x 1 mL	FXIII Inhibitor Reagent	2305100/4	2019-04-30
1 x 6 mL	Stabilizer solution	2305100/5	2019-04-30

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B00	58.2 – 97.0 % 83.8
	Coagulation Control A Lot: 3P71B00	28.8 – 48.0 % 41.4

Durchführung von
Prepared by

Michael KAFKA 

03. MAI 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER 

03. MAI 2018

Datum / Date: