
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRR00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601
LOT RR83B00.01
 2019-08-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 Ag Plate	RRS81B0	2020-03-31
1 x 0.3mL	Adamts13 Ag Conjugate	RQV71B0	2019-11-30
1x 80mL	Wash buffer Concentrate	RNA81B0	2020-01-31
1 x 90mL	Incubation buffer	RNB81B0	2020-02-29
1 x 12mL	TMB Substrate	RKN81B0	2020-01-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag High Control	RRY81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Low Control	RRZ81B0	2020-02-29

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.95	OK
Verwendbarkeit Usability	High Control (RRY81B0)	0.53 – 0.89 IU/mL	0.68	OK
	Low Control (RRZ81B0)	0.12 – 0.28 IU/mL	0.22	OK

**Durchführung von
Prepared by**

Michael KAFKA 

19. APR. 2018

Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

19. APR. 2018

Datum / Date: