



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ7UT0AZ.02

TECHNOZYM[®] ADAMTS-13 Activity Control Set 2x0.5mL

REF 5450763

LOT 7UT81B0.01

 2019-06-30

Inhalt Content	Reagenz Reagent	LOT	
1x0.5mL	Adamts-13 Activity High Control	RUY71B0	2019-06-30
1x0.5mL	Adamts-13 Activity Low Control	RUZ71B0	2019-06-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit / Usability	High Control (RUY71B0)	0.59 – 0.98 IU/mL	0.71	OK
	Low Control (RUZ71B0)	0.13 – 0.31 IU/mL	0.21	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

22. MAI 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER *Eva Maria Hager*

22. MAI 2018

Datum / Date: