

**CERTIFICATE OF ANALYSIS**

***CHROMOGENIX***

PRODUCT:	S-2238, 25 mg Substrate	PRODUCT #:	82032439
EXP. DATE:	2021-01	LOT #:	N0184814
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	820324

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2238	N0184814	2021-01	6 ml

PREPARED BY: Maia Dettinger DATE: 1-18-18

APPROVAL SIGNATURE: P. Pencza DATE: 1/18/18