

**CERTIFICATE OF ANALYSIS**



PRODUCT NAME :	<u>COAMATIC FACTOR VIII</u>	EXPIRATION DATE: <u>2020-02</u>
PRODUCT NUMBER :	<u>82258563</u>	
LOT NUMBER:	<u>N0386233</u>	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.43	PASS
Intercept	---	0.12	--	0.04	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.6	PASS
Intercept	---	0.21	--	0.08	PASS
r	0.990	---	--	1.000	PASS

We hereby certify that the above product(s):  
 · have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;  
 · meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: N. Patel *N.M. Patel* DATE: 3/30/2018  
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: D. Tencza *D. Tencza* DATE: 3/30/2018  
 PRINT SIGNATURE

TITLE: Senior QC Analyst