

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2020-03
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0386240		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Coatest SP4 Factor VIII Kit (82409463)					
High Range					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.114	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.038	PASS
Intercept	--	0.406	--	0.154	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS

PREPARED BY: Nehal Patel *N.m. Patel*
 PRINT SIGNATURE

DATE: 4/5/2018 *4-5-18*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
 PRINT SIGNATURE

DATE: 4/5/2018 *4-5-18*

TITLE: Senior QC Analyst