

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2021-03
PRODUCT NUMBER:	82141339		
LOT:	N0386214		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.0	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Marie Dettlinger Marie Dettlinger DATE: 3-21-18
 PRINT SIGNATURE

TITLE: QC Analyst

APPROVAL SIGNATURE: Nehal Patel N.M. Patel DATE: 3-21-18
 PRINT SIGNATURE

TITLE: Senior QC Analyst