

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ9K00AZ
---	---	----------

## Ceveron® TGA BUF 5x1mL

REF      5006370  
LOT      9K81B.02  
      2020-02-29

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
5x1 mL	Ceveron® TGA BUF	9K81B	2020-02-29

Analytische Prozedur Analytical Procedure			Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls <span style="border: 1px solid black; padding: 2px;">LOT</span> 1Y71B00	Peak Thrombin nM	Ceveron® TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D82B	215 – 358	232.4	OK
	Lag time min	Ceveron® TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D82B	2.5 – 4.1	3.1	OK
	AUC nM	Ceveron® TGA RC Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 7D82B	1975 - 3291	2636	OK
Etiketten und Barcode Überprüfung Label and bar code Verification				OK	

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

11. JUNI 2018

Datum / Date:

Freigabe von  
Approved by

**Alessandro BENIGNI**

*11-Jun-2018*

Datum / Date: