



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0H00.01

TECHNOVIEW LMW CAL Set 5x1mL

REF 5090040

LOT 0H81B00.01

🕒 2020-03-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1mL	TECHNOVIEW LMW Cal 1	5A81B00	2020-03-31
1x1mL	TECHNOVIEW LMW Cal 2	5B81B00	2020-03-31
1x1mL	TECHNOVIEW LMW Cal 3	5C81B00	2020-03-31
1x1mL	TECHNOVIEW LMW Cal 4	5D81B00	2020-03-31
1x1mL	TECHNOVIEW LMW Cal 5	5E81B00	2020-03-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW LMW Cont. L LOT 3F81B00	0.25 – 0.47 IU/mL	0.39	OK
	TECHNOVIEW LMW Cont. H LOT 3H81B00	0.83 – 1.39 IU/mL	1.06	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von
Prepared by

Suzana GRESAK PODAKAR

27. APR. 2018

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

27. APR. 2018

Datum / Date: