



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0900AZ.01

TECHNOVIEW® Rivaroxaban Cal Set 5x1mL

REF 5090170

LOT 0981B00.01

⌚ 2019-11-30

| Inhalt<br>Content | Reagenz<br>Reagent            | LOT     | ⌚          |
|-------------------|-------------------------------|---------|------------|
| 1x1 mL            | TECHNOVIEW® Rivaroxaban Cal 1 | 5172B00 | 2019-11-30 |
| 1x1 mL            | TECHNOVIEW® Rivaroxaban Cal 2 | 5671B00 | 2019-11-30 |
| 1x1 mL            | TECHNOVIEW® Rivaroxaban Cal 3 | 5771B00 | 2019-11-30 |
| 1x1 mL            | TECHNOVIEW® Rivaroxaban Cal 4 | 5871B00 | 2019-11-30 |
| 1x1 mL            | TECHNOVIEW® Rivaroxaban Cal 5 | 5971B00 | 2019-11-30 |

| Analytische Prozedur<br>Analytical Procedure                         |                                   | Anforderung<br>Requirement | Ergebnis<br>Result |    |
|--|-----------------------------------|----------------------------|--------------------|----|
| Verwendbarkeit<br>Usability<br>Anti-Xa Rivaroxaban<br>Ceveron® alpha | Rivaroxaban Cont M<br>LOT 3551B01 | 114.9 – 191.5 ng/mL        | 151.1              | OK |
|  | Rivaroxaban Cont L<br>LOT 3472B00 | 40.8 – 75.8 ng/mL          | 61.7               | OK |
| Etiketten und Barcode Überprüfung<br>Label and bar code Verification |                                   |                            | OK                 |    |

Durchführung von  
Prepared by

Christoph ENNE

29. JAN. 2018

Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

31. JAN. 2018

Datum / Date: