

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00.01
---	---	-----------

TECHNOCHROM® C1-Inhibitor 30T.

REF 5345003
LOT 0J83B00.01
 2019-10-31

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N for C1-INH	1P72BC1K	2020-01-31
1x1mL	Coagulation Control A for C1-INH	3P73BC1K	2020-01-31
1x0.882 mL	Coagulation Reference for C1-INH	1R73BC1H	2019-10-31
1x3 mL	C1 Esterase	7B73B00	2019-12-31
1x3 mL	Substrate f. C1-Esterase	8C72B00	2022-06-30
1x25 mL	C1-INH Probenpuffer A	9M71B00	2022-10-31
1x20 mL	C1-INH Reaktionspuffer B	9N71B00	2022-10-31


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P72BC1K	88.1 – 146.9%	93.6	OK
	CCA LOT 3P73BC1K	41.7 – 69.5%	62.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

01. JUNI 2018
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

01. JUNI 2018
Datum / Date: