




**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451
LOT RP85B0S.01
 2019-04-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 6	Adamts13 INH Plate	RPS74BS	2019-04-30
1 x 0.3mL	Adamts13 INH Conjugate	RPV72B0	2019-21-31
1x 80mL	Wash buffer Concentrate	RNA81B0	2020-01-31
1 x 90mL	Incubation buffer	RNB81B0	2020-02-29
1 x 12mL	TMB Substrate	RKN81B0	2020-01-31
1 x 12mL	Stop solution	RKK72B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.33	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	90.09	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	25.97	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. MAI 2018

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

09. MAI 2018

Datum / Date: