

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6PT00AZ
---	---	-----------

**TECHNOZYM® ADAMTS-13 INH Calibrator Set 5x0.5mL**

REF      5450461  
LOT      6PT72B0.02  
      2019-08-31

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.63	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	84.95	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	27.21	OK

**Durchführung von  
Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. MAI 2018  
Datum / Date:

**Freigabe von  
Approved by**

Michael KAFKA *[Signature]*

09. MAI 2018  
Datum / Date: