

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQRU00.04
		Alt.05
		RU85B00.01

## TECHNOZYM® ADAMTS-13 Activity ELISA 96T.

REF	5450701
LOT	RU85B00.01
	2019-06-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
2x	6mL	ADAMTS-13 Activity vWF73 Substrate	820	2019-11-30
1x	30mL	ADAMTS-13 Activity Reaction Buffer	820	2019-11-30
1x	12mL	ADAMTS-13 Activity HRP Conj Antibody	820	2019-11-30
1x	12mL	ADAMTS-13 Activity Color Reagent	820	2019-11-30
1x	53mL	ADAMTS-13 Activity Wash Buffer Conc	820	2019-11-30
1x	12mL	ADAMTS-13 Activity Stop Solution	820	2019-11-30
1x	n.a.	ADAMTS-13 Activity Plate 96T.	820	2019-11-30
1x	n.a.	ADAMTS-13 Activity sample dilution MTP	820	2019-11-30
3x	n.a.	ADAMTS-13 Activity Plate Seal	820	2019-11-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 1	1UT71B0	2019-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 2	2UT71B0	2019-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 3	3UT71B0	2019-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 4	4UT71B0	2019-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 6	6UT71B0	2019-06-30
1x	0.5mL	UY ADAMTS-13 Activity High Control	RUY71B0	2019-06-30
1x	0.5mL	UZ ADAMTS-13 Activity Low Control	RUZ71B0	2019-06-30
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 5	5UT71B0	2019-06-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal6	≥ 0.7	1.4	OK
Kontrollwiederfindung Recovery of controls	ADAMTS-13 Activity High Control LOT RUY71B0	0.59 - 0.98 IU/mL	0.74 IU/mL	OK
	ADAMTS-13 Activity Low Control LOT RUZ71B0	0.13 - 0.31 IU/mL	0.23 IU/mL	OK



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQRU00.04

Att.05

RU85B00.01


Durchgeföhrt von  
Prepared by

Michael KAFKA 

13. JUNI 2018

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

13. JUNI 2018

Datum / Date: