

# CERTIFICATE OF ANALYSIS

## CHROMOGENIX

Instrumentation  
Laboratory

A Werfen Company

PRODUCT NAME:	<u>CoaTest SP Factor FVIII Kit</u>	EXPIRATION DATE:	<u>2020-04</u>
PRODUCT NUMBER:	<u>82408663</u>		
LOT NUMBER :	<u>N0487397</u>		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0084	PASS
Intercept	--	0.250	--	0.106	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.042	PASS
Intercept	--	0.406	--	0.140	PASS
Standard Curve (r)	0.99	--	--	1.000	PASS

PREPARED BY:     *Maria Cettinger M. Stutzen*     DATE:     5-16-18      
 PRINT SIGNATURE

TITLE     QC Analyst    

APPROVAL SIGNATURE:     *Donna Tencza*     DATE:     5/16/18      
 PRINT SIGNATURE

TITLE     Senior QC Analyst