
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7F00AZ
---	---	----------

## TECHNOTHROMBIN® TGA RC High 50x0.5mL

REF      5006216

LOT      7F72B00.02

      2019-08-31

Inhalt Content	Reagenz Reagent	LOT	
50x0.5 mL	TGA RC High	7F72B00	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z81B00	420 - 700	576.9	OK
	TGA Control Low (nM Peak Thrombin) LOT 2881B00	257 - 429	324.1	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

23. APR. 2018

Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *[Signature]*

23. APR. 2018

Datum / Date: