
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 5x1.5mL

REF 5006235

LOT 8F63B01.01

 2020-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x1.5mL	TGA Substrat	8F63B01	2020-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z81B00	TGA RC Low LOT 7D81B00	257 - 429	376.4 OK
		TGA RD LOT 7G71B00	476 - 793	631.6 OK
	TGA Control Low (nM Peak Thrombin) LOT 2881B00	TGA RC Low LOT 7D81B00	132 - 220	190.4 OK
		TGA RD LOT 7G71B00	454 - 756	665.9 OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

29. JUNI 2018

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

02. JULI 2018

Datum / Date: