



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0T00AZ

TECHNOTHROMBIN<sup>®</sup> TGA Kit 3x16 T.


**REF** 5006010

**LOT** 0T82B00.02

2018-12-31

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x3mL	TGA buffer	9P31B03	2019-11-30
3x1.5mL	TGA substrat	8F71B00	2019-04-30
1x1.5mL	TGA reagent RD	7G61B01	2019-01-31
1x0.5mL	TGA reagent RC Low	7D62B01	2019-01-31
1x0.5mL	TGA reagent RC High	7F72B00	2019-08-31
1x0.5mL	TGA Thrombin Calibrator	5T71B	2019-07-31
1x1mL	TGA Control High	1Z72B00	2018-12-31
1x1mL	TGA Control Low	2881B00	2019-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung RC High	TGA Control High (nM Peak Thrombin) <b>LOT</b> 1Z72B00	375 - 625	468.3	OK
Recovery of controls RC High	TGA Control Low (nM Peak Thrombin) <b>LOT</b> 2881B00	257 - 429	273.3	OK
Kontrollwiederfindung RC Low	TGA Control High (nM Peak Thrombin) <b>LOT</b> 1Z72B00	251 - 418	271.3	OK
Recovery of controls RC Low	TGA Control Low (nM Peak Thrombin) <b>LOT</b> 2881B00	132 - 220	207.0	OK
Kontrollwiederfindung RD	TGA Control High (nM Peak Thrombin) <b>LOT</b> 1Z72B00	432 - 720	562.6	OK
Recovery of controls RD	TGA Control Low (nM Peak Thrombin) <b>LOT</b> 2881B00	454 - 756	530.9	OK

	<p style="text-align: center;">ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</p>	<p style="text-align: center;">DQ0T00AZ</p>
---	---	---

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

02. FEB. 2018  
Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

02. FEB. 2018  
Datum / Date: