

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00.01
---	---	-----------

Coagulation Reference, 5x1mL

REF 5220110
LOT 1R71B01.01
 2020-01-31


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Factor XII, CCN LOT 1P71B00	89.3 – 148.9 %	103.3	OK
	Factor XII, CCA LOT 3P71B00	43.1 – 71.8 %	45.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lavdrim NASUFI

16. APR. 2018
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

17. APR. 2018
Datum / Date: