
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRP00AZ
---	---	----------

**TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T**

REF 5450401  
LOT RP87B00.01  
 2019-08-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS85B0	2020-05-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV81B0	2020-04-30
1x 80mL	Wash buffer Concentrate	RNA81B0	2020-01-31
1 x 90mL	Incubation buffer	RNB81B0	2020-02-29
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT71B0	2019-08-31
1 x 0.5mL	Adamts13 INH High Control	RPY71B0	2019-08-31
1 x 0.5mL	Adamts13 INH Low Control	RPZ71B0	2019-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.86	OK
	Calibrator 5	< 0.1	0.06	OK
Verwendbarkeit Usability	High Control (RPY71B0)	60.24 – 100.40 U/mL	83.15	OK
	Low Control (RPZ71B0)	16.64 – 38.84 U/mL	27.38	OK

**Durchführung von  
Prepared by**

*Silwet Naschi*  
Silwet NASUFI

22. JUNI 2018

Datum / Date:

**Freigabe von  
Approved by**

*Christoph Enne*  
Christoph ENNE

22. JUNI 2018

Datum / Date: