



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM[®] ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR84B00.01

🕒 2019-11-30

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 12	Adamts13 Ag Plate	RRS83B0	2020-06-30
1 x 0.3mL	Adamts13 Ag Conjugate	RQV71B0	2019-11-30
1x 80mL	Wash buffer Concentrate	RNA81B0	2020-01-31
1 x 90mL	Incubation buffer	RNB81B0	2020-02-29
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag High Control	RRY81B0	2020-02-29
1 x 0.5mL	Adamts13 Ag Low Control	RRZ81B0	2020-02-29

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	1.19	OK
Verwendbarkeit Usability	High Control (RRY81B0)	0.53 – 0.89 IU/mL	0.62	OK
	Low Control (RRZ81B0)	0.12 – 0.28 IU/mL	0.19	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

04. JULI 2018
Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

04. JULI 2018
Datum / Date: