

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRXXAZ
---	---	---------

TECHNOZYM® t-PA Combi Actibind ELISA Kit 96T

REF TC16000
LOT 160001B8.01
 2019-03-31

Inhalt Content	Reagenz Reagent	LOT	
1x	anti t-PA Plate	RGN81B0	2020-03-31
1x0.3mL	POX anti t-PA	RKM61B1	2019-12-31
1x80mL	Wash buffer Concentrate	RNA81B0	2020-01-31
1x90mL	Incubation buffer	RNB81B0	2020-02-29
1x12mL	TMB Substrat	RKN81B0	2020-01-31
1x12mL	Stop solution	RKK72B0	2019-08-31
1x0.5mL	Standard t-PA	RAW11B4	2020-02-29
1x20mL	Detection mix buffer- TRIS	RDA71B0	2019-03-31
1x	Detection mixture	RBA71B0	2019-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1 activity	> 0.7	1.01	OK
	Calibrator 5 activity	< 0.1	0.05	OK
	Calibrator 1 antigen	> 1.0	1.66	OK
	Calibrator 5 antigen	< 0.1	0.00	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

11. APR. 2018

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *[Signature]*

11. APR. 2018

Datum / Date: