
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ6S0B.02
---	---	-----------

Siron LS (aPTT liquid) 10x10 mL

REF 5035109
LOT 6S76BB0.01
 2019-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
pH		7.0 – 7.6	7.1	OK	
Sterilität sterility		0 CFU/mL	0	OK	
Sensitivität Lupus sensitivity lupus	Lupus Inhibitor Plasma	≥ 1.5	3.7	OK	
aPTT Verwendbarkeit aPTT Usability Ceveron® alpha		CCN LOT 1P71B00	23.0 – 28.4 sec	25.7	OK
		CCA LOT 3P71B00	46.8 – 57.8 sec	56.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK		

Durchführung von
Prepared by


Eva Maria HAGER

15. MRZ. 2018
Datum / Date:

Freigabe von
Approved by


Stefan STEINER

16. MRZ. 2018
Datum / Date: