

	<p style="text-align: center;"><b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b></p>	<p style="text-align: right;">DQ0Z00AZ.02</p>
---	--	---

## TECHNOCHROM<sup>®</sup> FXIII 3 x 3 mL

REF      5360010  
LOT      0Z84B00.01  
      2019-06-30

Inhalt	Reagenz	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
3 x 3 mL	FXIII Activator Reagent	2324100/1	2019-06-30
3 x 3 mL	FXIII Detection Reagent	2324100/2	2019-06-30
3 x 3 mL	FXIII NADPH	2324100/3	2019-06-30
3 x 1 mL	FXIII Inhibitor Reagent	2324100/4	2019-06-30
1 x 6 mL	Stabilizer solution	2324100/5	2019-06-30

Test	Spezifikation	Ergebnis	
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B01	58.2 – 97.0 %	75.9
	Coagulation Control A Lot: 3P71B01	28.8 – 48.0 %	31.0

Durchführung von  
Prepared by

  
**Eva Maria HAGER**

**16. AUG. 2018**

Datum / Date:

Freigabe von  
Approved by

  
**Stefan STEINER**

**16. AUG. 2018**

Datum / Date: