


**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**


**Instrumentation  
Laboratory**  
 A Werfen Company

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2020-10</u>
PRODUCT NUMBER:	<u>82098539</u>		
LOT NUMBER:	<u>N0487240</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>FACTOR Xa, 71 nkat (650515)</b>					
ACTIVITY	61	81	nkat/vial	69	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	67.8	75.0	nkat/vial	71.8	PASS

PREPARED BY: Maria Oettinger   
 PRINT SIGNATURE

DATE: 5/24/2018 5-24-18

TITLE: QC Analyst

APPROVAL SIGNATURE: Nehal Patel   
 PRINT SIGNATURE

DATE: 5/24/2018 5-24-18

TITLE: Senior QC Analyst