

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2020-05
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0688436	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.45	PASS
Intercept	---	0.12	--	0.05	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.8	PASS
Intercept	---	0.21	--	0.10	PASS
r	0.990	---	--	1.000	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N. M. Patel*
 PRINT SIGNATURE

DATE: 6-26-18
6/26/2018

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*
 PRINT SIGNATURE

DATE: 6/26/18
6/26/2018

TITLE: Technical Supervisor