

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2021-04
PRODUCT NUMBER:	82141339		
LOT:	N0487245		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 5/8/2018 *5-8-18*
 PRINT SIGNATURE

TITLE: QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 5/8/2018 *5/8/18*
 PRINT SIGNATURE

TITLE: Senior QC Analyst