

CERTIFICATE OF ANALYSIS

CHROMOGENIX

PRODUCT:	S-2403, 25 mg Substrate	PRODUCT #:	82225439
EXP. DATE:	2020-11	LOT #:	N1173749
VENDOR:	INSTRUMENTATION LABORATORY	ASSAYED PER QCM:	822254

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.2	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

DESCRIPTION	LOT NUMBER	EXP. DATE	VOLUME
S-2403	N1173749	2020-11	3 ml

PREPARED BY: *Maria Dettner*

DATE: 11-29-17

APPROVAL SIGNATURE: *P. Ancoya*

DATE: 11/30/17