



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ7R00AZ

Ceveron<sup>®</sup> TGA RC Low 5x1 mL

REF 5006224  
LOT 7R81B.01  
⌚ 2020-03-31

<b>Inhalt Content</b>	<b>Reagenz Reagent</b>	<b>LOT</b>	<b>⌚</b>
5x1 mL	Ceveron <sup>®</sup> TGA RC Low	7R81B	2020-03-31

<b>Analytische Prozedur Analytical Procedure</b>		<b>Anforderung Requirement</b>	<b>Ergebnis Result</b>	
Kontrollwiederfindung Recovery of controls	TGA Control Low Lag time (min) LOT 2781B00	5.5 – 9.1	6.1	OK
	TGA Control Low Peak Thrombin (nM) LOT 2781B00	56 – 93	89.1	OK
	TGA Control Low AUC LOT 2781B00	665 – 1108	987.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Michael KAFKA

30. JULI 2018

Datum / Date:

Freigabe von  
Approved by

Eva Maria HAGER

30. JULI 2018

Datum / Date: