



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3F00.01

TECHNOVIEW® LMW Low CONT 5x1mL

REF 5090042

LOT 3F81B00.01

⌚ 2020-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.25 – 0.47 IU/mL	0.36	OK
Verwendbarkeit Usability Ceveron® Alpha	Anti-Xa LMWH	0.25 – 0.47 IU/mL	0.41	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Christoph ENNE

22. JUNI 2018

Datum / Date:

Freigabe von
Approved by


Laydrin NASUPI

22. JUNI 2018

Datum / Date: