



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3H00.02

TECHNOVIEW® LMW CONT H 5x1ml

REF 5090046

LOT 3H81B00.01

🕒 2020-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti-Xa	0.83 – 1.39 IU/mL		
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa	0.83 – 1.39 IU/mL	1.11	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Christoph ENNE

22. JUNI 2018

Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

22. JUNI 2018

Datum / Date: