



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0900AZ.01

TECHNOVIEW® Rivaroxaban Cal Set 5x1mL

REF 5090170

LOT 0982B00.01

⌚ 2020-07-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 1	5183B00	2020-07-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 2	5681B00	2020-07-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 3	5781B00	2020-07-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 4	5881B00	2020-07-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 5	5981B00	2020-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Anti-Xa Rivaroxaban Ceveron® alpha	Rivaroxaban Cont M LOT 3551B01	114.9 – 191.5 ng/mL	151.1	OK
	Rivaroxaban Cont L LOT 3472B00	40.8 – 75.8 ng/mL	52.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Christoph ENNE

11. SEP. 2018

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

11. SEP. 2018

Datum / Date: